

Serial No. C 106904

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

SUNNINGDALE LTD
COMMERCIAL ROAD
BLACKBURN
LANCS
BB30DS
LOCATION
IN/OUT

Machine Details

Cust ref	133904
Make	BILANCIAI
Model	FLUNT 50
Ser No	T232303022
Capacity	50000kg
e/d	20/10
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
1-20	T11077	17-1-23

Cert Interval	ANNUAL
Next Cert	JULY 2024

Tolerance +/-	KL
0-10000	20
10000-20000	40
20000-30000	60

Comments: ALL WEIGHTS IN UK
FORM LIFT - 4800
TEST UNIT - 21740
NEW VERIFICATION APPROX
INDICATION CHANGE

2	9	6
1	3	5

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
400	400	0	400	0
10000	10000	0	10000	0
19000	19000	0	19000	0
25740	25740	0	25740	0
29740	29740	0	29740	0
35740	35740	0	35740	0
40740	40740	0	40740	0
45540	45540	0	45540	0

Repeatability

Load	Ind	Ind	Ind

Repeatability

Load	Ind	Ind	Ind
45540	45540	45540	45540

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Eccentric Load - 10000

Reading	Dev	Reading	Dev
1 10000	0	2 10000	0
3 10000	0	4 10000	0
5 10000	0	6 10000	0
7		8	

Signature

Name CHAIL HICSEN

Date 24-7-23