



FM584346
ISO 9001:2008
QUALITY
MANAGEMENT
SYSTEMS



Members of the
UK Weighing Federation
Registration No. 2154099



Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com

CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

SHIRLEYS RECYCLING LTD
COMMERCIAL ROAD
BLACKBURN INTERCHANGE
DARWEN
LOCATION BB3 0DB.
YARD.

Machine Details

Cust ref	
Make	BILANCIAI
Model	D800
Ser No	T10110631
Capacity	50 / 000 / Kg
e/d	1 20 / Kg
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
N21	W16015	22/4/16

Cert Interval	12 MONTH
Next Cert	JULY 2017

Tolerance +/-	Kg
0- 10000	20
10000-40000	40
40000-50000	60

Comments ALL READINGS IN Kg
.....
.....
.....

2	4	6
1	3	4
x		

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
2000	2000	0	0	0
8000	8000	0	2000	0
12000	12000	0	8000	0
16000	16000	0	12000	0
20000	20000	0	16000	0
21580	21580	0	21580	0
31580	31580	0	31580	0
41580	41580	0	41580	0
46440	46420	-20	46420	-20

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
46420	46420	46420	46420

Repeatability

Load	Ind	Ind	Ind

Eccentric Load

Reading	Dev	Reading	Dev
1 6860	0	2 6860	0
3 6860	0	4 6860	0
5 6860	0	6 6860	0
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature

Name S. FILLINGHAM

Date 7/7/16